

Issued by NMI Certin B.V.,
designated and notified by the Netherlands to perform tasks with respect to
conformity assessment procedures mentioned in Article 13 of Directive
2014/31/EU, after having established that the measuring instrument meets
the applicable requirements of Directive 2014/31/EU, to:

Manufacturer Rinstrum Pty Ltd
Unit 4/31 Henry Street
Loganholme, Queensland, 4129
Australia

Measuring instrument **A Non-automatic weighing instrument**
Type : R3xx/R4xx/C5xx/X3xx/C3xx

Further properties are described in the annex:
– Description T12994 revision 0.

Valid until 13 February 2035

Initially issued 13 February 2025

Issuing Authority **NMI Certin B.V., Notified Body number 0122**
13 February 2025

Certification Board

This document is issued under the provision that
no liability is accepted and that the
manufacturer shall indemnify third-party
liability.

The designation of NMI Certin B.V. as Notified
Body can be verified at [http://
ec.europa.eu/growth/tools-databases/nando/](http://ec.europa.eu/growth/tools-databases/nando/)

Reproduction of the complete
document only is permitted.

This document is digitally signed
and sealed. The digital signature
can be verified in the blue ribbon at
the top of the electronic version of
this certificate.



Description

Number **T12994** revision 0
Project number 3897089
Page 1 of 2

1 General information about the non-automatic weighing instrument

All properties of the non-automatic weighing instrument, whether mentioned or not, shall not be in conflict with the legislation.

This certificate contains references to other certificates. The properties mentioned in these certificates shall be observed in addition to the properties mentioned in this certificate.

1.1 Essential parts

Indicator:

Producer	Type	Certificate number
Rinstrum Pty Ltd.	R3xx series	TC11437
Rinstrum Pty Ltd.	R4xx series	TC6821
Rinstrum Pty Ltd.	C5xx series, WE2111	TC11766
Rinstrum Pty Ltd.	X3xx	TC7419
Rinstrum Pty Ltd.	C3xx Series	TC12863

Any analog passive load cell(s) may be used under this certificate for instruments as described in WELMEC 2.4 Issue 2021, provided the following conditions are met:

- There is a respective certificate (EN45501) or an OIML Certificate (R 60) issued for the load cell by a Notified Body responsible for type examination under Directive 2014/31/EU or Directive 2014/32/EU;
- The load cell shall comply with humidity class **CH** or **SH**;
- The certificate contains the load cell types and the necessary load cell data required for the manufacturer's declaration of compatibility of modules (EN 45501:2015 clause F.4), and any particular installation requirements;
- The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in EN 45501:2015 clause F.4, at the time of putting into use;
- The load transmission must conform to one of the examples shown in WELMEC 2.4 Issue 2021.

1.2 Essential characteristics

Accuracy class	Ⓜ or ⓂⓂ
Maximum number of scale intervals	$n \leq$ number of scale intervals in the certificates

1.3 Essential shapes

The data plate is secured against removal by sealing or will be destroyed when removed.

1.4 Conditional parts

All type of the instruments may be equipped with the following parts that further process the measurement result without modification under the conditions stated in the table:

Part	Condition(s)	Reference document
Simple recipient printer	CE marking present	WELMEC 2.10 clause 3.1.3
Printer Data Storage Device	CE marking present and the part is certified to be connected to a weighing instrument by a Notified Body responsible for type examination under Directive 2014/31/EU.	WELMEC 2.10 clause 3.1.3

When equipped with the R3xx series, the instrument may be equipped with the parts below::

Point of Sale (POS)	CE marking present and the part is certified to be connected to a weighing instrument by a Notified Body responsible for type examination under Directive 2014/31/EU.	WELMEC 2.2
Electronic Funds Transfer equipment (EFT)	CE marking present and the EFT represents only the price total on the display.	-

the instrument is installed in a fixed position. A ring on the level indicator indicates when the maximum tilt is The non-automatic weighing instrument is fitted with a levelling device and a level indicator, unless exceeded.

2 Seals

To secure components that may not be dismantled or adjusted by the user, the non-automatic weighing instrument has to be secured in a suitable manner on the locations indicated in the certificates involved.

3 Conditions for conformity assessment

The marks, facilities for the marks and the inscriptions on the non-automatic weighing instrument fulfil the requirements of Directive 2014/31/EU Annex III clause 1.

The compatibility of load cells and indicator is established by the manufacturer by means of the compatibility of modules form, contained in EN 45501:2015 clause F.4, at the time of putting into use.